



SQ1 - SUPERHEAT TEST



Name Techno Air Gate (TAG)
Address 2311 Abdulsalam Amin St., 2nd District, 6th October, Giza, Egypt.
Email/Tel info@technoairgate.com

Site

JAMJOOM PHARMA

Date

Monday, May 26, 2025

Location / Sterilizer

Gaulin Homogenizer

Code

--

TEST

Cycle No

1

2

3

Average

Temperature of Steam (°C)

136.0

139.0

138.3

Temperature in Expansion Tube (°C)

100.2

99.9

100.3

Boiling point at local condition (°C)

99.5

99.5

99.5

Superheat °C

0.7

0.4

0.8

0.6

Requirements of EN 285

≤ 25°C