



SQ1 - SUPERHEAT TEST



Name Techno Air Gate (TAG)
Address 2311 Abdulsalam Amin St., 2nd District, 6th October, Giza, Egypt.
Email/Tel info@technoairgate.com

Site

JAMJOOM PHARMA

Date

Sunday, May 25, 2025

Location / Sterilizer

Water station WFI Tank

Code

--

TEST

Cycle No

1

2

3

Average

Temperature of Steam (°C)

139.6

139.0

139.3

Temperature in Expansion Tube (°C)

99.9

100.8

100.2

Boiling point at local condition (°C)

99.5

99.5

99.5

Superheat °C

0.4

1.3

0.7

0.8

Requirements of EN 285

≤ 25°C