



SQ1 - SUPERHEAT TEST



Name Techno Air Gate (TAG)
Address 2311 Abdulsalam Amin St., 2nd District, 6th October, Giza, Egypt.
Email/Tel info@technoairgate.com

Site

JAMJOOM PHARMA

Date

Sunday, May 25, 2025

Location / Sterilizer

FEDEGARI Autoclave

Code

--

TEST

Cycle No	1	2	3	Average
Temperature of Steam (°C)	140.9	141.0	141.3	
Temperature in Expansion Tube (°C)	100.2	100.1	100.6	
Boiling point at local condition (°C)	99.5	99.5	99.5	
Superheat °C	0.7	0.6	1.1	0.8

Requirements of EN 285

≤ 25°C