



## SQ1 - NON-CONDENSABLE GAS TEST



**Name** Techno Air Gate (TAG)  
**Address** 2311 Abdulsalam Amin St., 2nd District, 6th October, Giza, Egypt.  
**Email/Tel** info@technoairgate.com

**Site**

JAMJOOM PHARMA

**Date**

Sunday, May 25, 2025

**Location / Sterilizer**

FEDEGARI Autoclave

**Code**

--

### TEST

**Cycle No**

1

2

3

**Average**

**Volume of gas at start of test (ml)**

0.0

0.0

0.0

**Volume of gas at end of test (ml)**

0.3

1.0

0.5

**Volume of condensate at start of test (ml)**

0.0

0.0

0.0

**Volume of condensate at end of test (ml)**

100.0

100.0

100.0

**Non-condensable gas level (%)**

0.3%

1.0%

0.5%

0.6%

**Requirements of EN 285**

NMT 3.5%