



SQ1 - NON-CONDENSABLE GAS TEST



Name Techno Air Gate (TAG)
Address 2311 Abdulsalam Amin St., 2nd District, 6th October, Giza, Egypt.
Email/Tel info@technoairgate.com

Site JAMJOOM PHARMA

Date Monday, May 26, 2025

Location / Sterilizer Gaulin Homogenizer

Code --

TEST

Cycle No	1	2	3	Average
Volume of gas at start of test (ml)	0.0	0.0	0.0	
Volume of gas at end of test (ml)	0.7	0.6	0.8	
Volume of condensate at start of test (ml)	0.0	0.0	0.0	
Volume of condensate at end of test (ml)	100.0	100.0	100.0	
Non-condensable gas level (%)	0.7%	0.6%	0.8%	0.7%

Requirements of EN 285 NMT 3.5%